



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

Florida Pipette Calibration, LLC
2190 Deer Pointe Crossing, Bradenton, FL 32402

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Mechanical Calibration
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

Initial Accreditation Date:

August 24, 2020

Issue Date:

December 8, 2020

Expiration Date:

March 31, 2023

Accreditation No.:

78127

Certificate No.:

L20-741

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjilabs.com



Certificate of Accreditation: Supplement

Florida Pipette Calibrations, LLC.
21910 Deer Pointe Crossing, Bradenton, FL 34202
Contact Name: Debbie Martella Phone: 866-357-4738

Accreditation is granted to the facility to perform the following calibrations:

Mechanical Calibration

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Pipette and Bottle Top Dispenser ^{FO}	0.5 μ L to 2 μ L	0.08 μ L	Mettler-Toledo Precision Balance. FPCCAL ISO 101 ISO 8655
	2 μ L to 10 μ L	0.08 μ L	
	10 μ L to 20 μ L	0.08 μ L	
	20 μ L to 100 μ L	0.13 μ L	
	100 μ L to 200 μ L	0.19 μ L	
	200 μ L to 1 000 μ L	1.1 μ L	
	1 000 μ L to 5 000 μ L	3.1 μ L	
	5 000 μ L to 10 000 μ L	4.7 μ L	
	10 000 μ L to 25 000 μ L	21 μ L	
	25 000 μ L to 50 000 μ L	43 μ L	
Hamilton Diluter ^O	0.5 μ L to 25 μ L	0.07 μ L	
	25 μ L to 250 μ L	0.09 μ L	
	250 μ L to 2 500 μ L	0.17 μ L	

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript O means that the laboratory performs calibration of the indicated parameter onsite at customer locations. Example: Outside Micrometer^O would mean that the laboratory performs this calibration onsite at the customer's location.
4. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.